



## CERTIFICATE OF ANALYSIS (COA)

**VIROSTUFF™ - Viral Transport Medium**  
Vial specimen transport medium

Product Name: VIROSTUFF™	Catalog # PRX-VS01- 50T	Lot / Batch # PRX-VS-0921-001 PRX-VS-0122-004
Analysis Date: October 16, 2021	Expiration Date: PRX-VS-0921-001 @ 15 Oct 2022 PRX-VS-0122-004 @ 09 Jan 2023	Storage conditions: 2 TO 28° C / Room temperature

### Product description

Viral transport media are prepared with the idea of sustaining the viability of the viral culture or specimens for tests like the nucleic acid amplification test (NAAT) while preventing the drying of the sample. The Virostuff is a Viral Transport Medium (VTM) allows the safe transfer of viruses for further research, including conventional cell culture methods, diagnostic tests, and molecular biology techniques. VIROSTUFF VTM tubes are available in a screw cap plastic tube containing buffered proteins (serum, albumin or gelatin) and antibiotics. Antibiotics are usually incorporated into the viral transport media to suppress the growth of contaminating bacteria and fungi.

### Material Ingredients:

- Inactivated Fetal Bovine Serum (FBS)
- Hanks Balanced Salt Solution (HBSS).
- Antibiotics 100 µg/mL Gentamicin and 0.5 µg/mL Amphotericin B
- Aliquot 3mL of the prepared medium into individual sterile conical screw-capped tubes.
- Store at 2-28°C / Room Temperature.



**PRAWNRX**

47 A, Asambu Road,  
Vadasery, Nagarcovil,  
Kanyakumari District,  
Tamilnadu, India- 629001,

PRAWNRX HEALTHCARE Pvt. Ltd.

Email: prawnrx@gmail.com


Tel: +91-8754888652

Ph: 04652 - 295352

**Analysis:**

Quality Control Test	Specification	Result
Sterility (No visual turbidity)- upto 120 hrs	Clear, No Color change	Pass
Sterility (Bacterial) – Blood Agar Plate	No growth	Pass
Sterility (Yeast/Mold) – Peel Plate <sup>®</sup> Y/M	No growth	Pass
SARS-CoV-2 (Coronavirus) RNA-Free	Negative	Pass
SARS-CoV-2 (Coronavirus) spike samples	Positive	Pass
Viral Transport Medium Volume	3 ± 0.2 mL	Pass
Viral Transport Medium pH	8.3 ± 0.2	Pass
Vial Leak Check	No leakage	Pass

Certificate can be referred for Research or Validation purpose only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services @ 04652 295352 or email us at [prawnrx@gmail.com](mailto:prawnrx@gmail.com)

<p>PRAWN Rx Health Care Pvt. Ltd. 47-A, Asambu Road, Vadasery, Nagercoil, Kanyakumari District, TamilNadu, India- 629001 Tel: +91- 4652 295352 Email: <a href="mailto:prawnrx@gmail.com">prawnrx@gmail.com</a></p>	 ----- Dr. K.P. Srinivasakumar QA/ QC In charge
--	---

Version 2.0 Issue Date: 10 January 2021

